

Ordering Information

Each and every order for a **GROM** chromatography column includes not only the desired hardware, but also

a Test Chromatogramm from that individual column produced in the **GROM** Quality Insurance Laboratory

a Quality Status Report to ensure customer feedback conform with GLP and ISO 9001 Standards

and, last not least our newest product, i.e., a package of "**GROM** Gummi Bears"
(subject to local purchasing agreements and federal import restrictions; i.e. not available in USA)



Our new ordering system: - Quick, simple and unequivocal -

In this new catalogue, a large number of stationary phases have been included for the first time, among others more than 20 new phases from our own range of packing materials. This would have placed great strains on our old order number system; the inclusion of a further single stationary phase in the existing palette could for instance mean that over 5500 new order numbers would have had to have been introduced. For this reason and in anticipation of

the continually growing demands in HPLC, changes in this order number system were unavoidable.


A modern order number code must satisfy primarily the following requirements:

- It should be self-explanatory and understandable for the customer
- It should be open for future developments, i.e., there should be no numerical limitation.
- It should be computer-compatible

In order to satisfy these demands, we must accept a minor blemish - up to 15 positions may be necessary to fully define an article like a common HPLC column, although a typical order number will have only 14 positions as described below. Nevertheless the new system is simple and easy to use, since it includes all parameters required to specify an HPLC column.

The new Order Numbers include all parameters required to specify an HPLC column

Tradename/ manufacturer	Basic material	Phase coating, specialty columns	Phase type	Particle diameter (µm)	Pore diameter (nm)*	Hardware type	Column length (cm)	column diameter (mm, resp. µm x 0.1**)
----------------------------	-------------------	-------------------------------------	------------	---------------------------	------------------------	------------------	-----------------------	---



Example
To order the following HPLC column:

GROM-SIL, ODS 4, Particle diameter 3 µm, Pore diameter 10 nm*, C(K)artridge 25 cm long, 4 mm i.d. or µm x 0.1*****

we abbreviate:



G S O D 4 0 3 1 0 K 2 5 0 4

This means:

All parameters required to describe an HPLC column are included in the new order number. Using the abbreviations shown on the next page, the ordering number is quickly and easily established, and speedy and correct processing of your order is guaranteed.

* 10nm = 100Å

** English „cartridge“ = German „Kartusche“

*** for capillary columns „µm x 0.1“

Order number abbreviations and letter code at the following positions:

1st. position of the order number

Abbreviation/Tradename	Manufacturer
B	μ Bondapak Waters (Milford, MA, USA)
C	Capcell Shiseido (Tokyo, Japan)
E	Eurospher Eurochrom-Knauer (Berlin, FRG)
G	GROM-Sil Grom (Herrenberg, FRG)
H	Hypersil Shandon (Runcorn, UK)
I	Inertsil Gl-Science (Tokyo, Japan)
K	Kromasil Eka-Nobel (Göteborg, Sweden)
L	Synchropak Synchrom (Lafayette, NJ, USA)
M	Lichrosorb Merck (Darmstadt, FRG)
M	Lichrospher Merck (Darmstadt, FRG)
M	Superspher Merck (Darmstadt, FRG)
N	Nucleosil u. Polygosil Macherey-Nagel (Düren, FRG)
P	Partisil Whatman (Clifton, NJ, USA)
R	PRP-Phases Hamilton (Reno, NV, USA)
S	Spherisorb PhaseSep (Deeside, UK)
V	Vydac Vydac (Hesperia, CA USA)
Z	Zorbax and Ultron Rockland Techn. Inc., (Rockland, DE, USA)

2nd. position of the order number

Abbreviation	Basic-material
S	Silica gel
P	Polymeric
A	Aluminium oxide
T	Titanium oxide
C	Calcium hydroxide
K	Graphitized carbon (German „Kohlenstoff“)

3rd. und 4th. position of the order number

Abbreviation	Phases and specialty columns
Stationary phases:	
AC	Amino-Cyano-phase
BA	Basic (for bases, e.g., amines)
BU	Butyl- / C4-
CH	Chiral phases
CN	CN- / cyanopropyl-/nitrile-
DM	Dimethyl-, C2-phase
HE	Hexyl- / C6-
HI	Hydrophobic Interaction Chromatography/HIC/ salting-out chromatography
NH	NH ₂ - / Aminopropyl-
NO	Nitropropyl-
NP	Normal phase silica
OA	ODS-Amino-, mixed-phase
OC	Octyl- / C8-
OD	ODS / Octadecyl- / C18-
OH	OH- / Diol phase on silical
PH	Phenyl-
PR	Propyl-
SA	Strong Anion exchanger / SAX /
SC	Strong Cation exchanger / SCX /
SE	SEC / Steric Exclusion Chromatography / GPC /
TM	TMS / Trimethylsilane- / C 1-
WA	Weak Anion exchanger / WAX
WC	Weak Cation exchanger / WCX

3rd. and 4th. position of the order number

Abbreviation	Phases and specialty columns
Specialty columns:	
AA	Acid Analysis Column
AL	Column for separation of Alcohols
BI	Column for separation of Bitter substances in beer
BR	Bio Rev. phase; special reversed phase column for the separation of proteins, peptides and nucleotides
DA	DABS-Cl/Dimethylaminoazobenzene-sulfonyl chloride; for amino acid analysis
FL	FLEC-/Fluorenyl ethyl chloroformate; for D/L-amino acid analysis
FM	FMOC-/9-Fluorenylmethyloxycarbonyl chloride; for amino acid analysis
OP	OPA-/ortho-phthaldialdehyde; for amino acid analysis
PE	Column for the analysis of pesticides
PA	Column for the analysis of PAH's, (polycyclic aromatic hydrocarbons)
PO	Special column for analysis of polyamines and biogenic amines Columns for analysis of sugars (German „Zucker“):
ZC	Cation exchanger/ Ca ⁺⁺ form
ZH	Cation exchanger/ H ⁺ form
ZO	Anion exchanger / OH ⁻ form
ZP	Cation exchanger / Pb ⁺⁺ form
ME	Method-development-set (German „Methoden Entwicklung“)

5th. position of the order number

Abbreviation	Phase type
If a manufacturer produces more than one phase of the same type (e.g., several C18 or C8 phases), these are differentiated by consecutive numbering (see pages 144-151). GROM-SIL phases with the same ligands are numbered according to increasing hydrophobicity (see comparison of selectivity on pages 10 and 11) and are thus characterised by their order number.	

10th. position of the order number

Abbreviation	Hardware
B	BIAX-column
C	Capillary column (for i.d. „μm x 0.1“ instead of „mm“ in the 13th. position)
C....T	Capillary column with short guard column head (tiny / 15th. position)
C....L	Capillary column with long guard column head (long / 15th. position)
G	Glass column
K	Cartridge (German „Kartusche“)
M	Mixed hardware, i.e. wetted parts made from PEEK, holding devices, threads, etc. from stainless steel (pressure resistant up to 40 MPa)
P	PEEK column (all parts made off PEEK)
S	Standard column
T	Triple-set (three identical cartridges from the same batch)
U	Up-scaling-set
V....V	Guard column (German „Vorsäule“; pk. of 5)
R	Refill

International Distributors

Berger Analysen- u. Informationstechnik GesmbH

Eichendorffstr. 24
AT-4020 Linz



Osterreich

Tel + 43 (0)732 / 346 287 - Fax -346 392
E-mail www.berger@gmx.net

Sercolab BVBA

Vaartdijk 55
B-2800 Mechelen



Belgium

Tel + 32 (0)15 / 421 319 - Fax -421 612
E-mail ruythoorenm@innet.be

WATREX Praha, s. r. o.

Hostalkova 42
16900 Praha 6



Czech Republic

Tel + 420 (0)2 / 20 513 446 - Fax -20 513 448
E-mail watrex@watrex.com

N. Asteriadis S.A.

12 Solomou Street
P.O.Box 26140
100 22 Athen



Greece

Tel + 30 (0)1 / 3 840 010 - Fax -8 239 567
E-mail info@asteriadis.gr

TECHCOMP - (H.K.) Limited

505 Shu Hing Centre
13 Sheung Yuet Road
Kowloon Bay



Hong Kong

Tel + 852 / 2751 9488 - Fax -2751 9217

LAB -COMP Kft.

Postfach 147
H-1 364 Budapest



Magyarország (Hungary)

Tel + 36 (0)1 / 2 806 770 - Fax -2 806 358
E-mail labcomp@mail.datanet.hu

APNA SCIENTIFIC SUPPLIES PVT. LTD.

M-2A, Industrial Estate
Guindy, Chennai 600 032



India

Tel + 91 (0)44 / 2 340 963 - Fax -2 331 107
E-mail apna@md3.vsnl.net.in

Biobusiness Development Agency

C-76, 1st. Floor, Fateh Nagar
Jail Road
New Dehli - 110 018



India

Tel + 91 (0)11 / 5 596 820 - Fax -5 144 316
E-mail bio.multani@access.net.in

M. AKSHAYKUMAR & Co

18/6 Sethia Sadan,
Sewree Wadala Scheme Rd.
Matunga Mumbaj - 400 019



India

Tel + 91 (0)22 / 201 6490 - Fax -206 5334
E-mail sscorp@bom3.vsnl.net.in

ISRAEL SCIENTIFIC INSTRUMENTS Ltd.

32 Shaham Street
Amargad House
Kiryat Matalon
P.O.Box 7790 / Petah-Tikva 49 170



Israel

Tel + 972 (0)3 / 9 232 202 - Fax -9 229 750
E-mail gadi.zoldan@isil.co.il

Eurisco Diagnostica Srl.

Via A.Fogazzaro, 62
36073 Cornedo (VI)



Italy

Tel + 390 (0)445 / 953 094 - Fax -953 192

SERVICE T.L. s.r.l.

Via Monte Nevoso 6
20131 Milano



Italy

Tel + 390 (0)2 / 2 664 384 - Fax -2 663 402

ThermoQuest Italia S.p.A.

Strada Rivoltana
20090 Rodano / Milano



Italy

Tel + 390 (0)2 / 950 591 - Fax -95 059 237
E-mail Germana.Buzzoni@ceinstruments.it

A-JOO SCIENTIFIC Co., Ltd.

Dong Nam Bldg.
258, Yangjae-Dong
Seocho-Ku, Seoul



International Distributors

NAM YOUNG ScientificCo.Ltd.

5th floor, # 501 Oh Sung B/D,
545-4, Dogok-dong,
Kang Nam-Ku, Seoul



Korea

Tel + 82 (0)2 / 5 294 781 - Fax -5 294 783

CHROMSCIENCE SDN BHD

23 - 0 - 27, Jalan 13/2A
Taman Wilayah Selayang
68 100 Batu Caves, Kuala Lumpur



Malaysia

Tel + 60 (0)3 / 6 163 260 - Fax -6 163 262

LEACSA, S.A. de C.V.

MAZATLAN # 71-BIS
COL. CONDESA
06140 MEXICO, D.F.



Mexico

Tel + 52 / 5211 0404 - Fax -286 4977

E-mail leacsa@mpsnet.com.mx

ZETATECH, LDA

Urbanização de Santo André,6-7ºB
2675 Póvos de Santo Adrião



Portugal

Tel + 351 (0)62 / 880 334 - Fax -880 330

E-mail chust@mail.telepac.pt

KEMOMED d. o. o.

Planina 3
4 000 Kranj



Slovenia

Tel + 386 (0)64 / 351 510 - Fax -351 511

E-mail kemomed@siol.net

NTK KEMI HB

Ekeby Bruk 35
75 275 Uppsala



Sweden

Tel + 46 (0)18 / 512 230 - Fax 512 240

STAGROMA AG

Chr. Merian-Ring 31a
CH-4153 Reinach



Schweiz

Tel + 41 (0)61 / 717 8 717 - Fax -717 8
710

E-mail stagroma@swissonline.ch

Pentad Scientific Corp.

6F-2, No. 181, Sec 2
Changan East Road
Taipei



Taiwan - R.O.C.

Tel + 886 (0)2 / 27 766 931 -Fax -27 766
932

E-mail pentaptp.@ms12.hinet.net

GIZEM TIP VE

LABORATUVAR CIHAZLARI

GIZEM Ltd.

MEDICAL & LABORATORY EQUIPMENTS

19 Mayıs Mahallesi - Sakacı sokak

Eligül Apt. No:83/1

81 080 Kazasker - Istanbul



Türkiye

Tel + 90 (0)216 / 4 645 130 - Fax -4 645
135

E-mail gizem@superonline.com

JONESChromatography

Tir-y-Berth Industrial Estate, New Road

Hengoed

Mid Glamorgan, CF 82 8AU



United Kingdom

Tel + 44 (0)1443 / 816 991 - Fax -816 552

WATREX International Inc.

3, Vantage Drive

Pittsford NY 14 534 - 3 205

USA

Tel / Fax + 1 (0)716 / 2 488 915

marek.minarik@watrex.com



ANATECH INSTRUMENTSLTD

13 Langwa Street/Straat 13 Strijdom Park,

P.O.Box 98485 - Sloane Park,

2152, Gauteng



South Africa (ZA)

Tel + 27 (0)11 / 7 923 300 - Fax -7 923 363